

Nome commerciale: **TERMOMETRO INFRAROSSI NON TOUCH**

Codice Prodotto: **TER-NTS**

Specification

Temperature units:

Celsius temperature($^{\circ}\text{C}$)/($^{\circ}\text{F}$), can be set

Measuring range:

Human body: ($32.0\sim 42.9^{\circ}\text{C}$)/($89.6\sim 109.2$) $^{\circ}\text{F}$

Object: ($0\sim 118$) $^{\circ}\text{C}$ /($32.\sim 244.4$) $^{\circ}\text{F}$

Accuracy: Human body: ($35\sim 42$) $^{\circ}\text{C}$

Outside $\pm 0.3^{\circ}\text{C}/\pm 0.54^{\circ}\text{F}$ ($35\sim 42$) $^{\circ}\text{C}$ Within $\pm 0.2^{\circ}\text{C}/\pm 0.4^{\circ}\text{F}$

Object: $\pm 1.0^{\circ}\text{C}/\pm 1.9^{\circ}\text{F}$



Forehead



One second
measurement



Memory
storage



Automatic
shutdown



EC Certificate
Directive 93/42/EEC Annex II, excluding Section 4
Full Quality Assurance System
Medical Devices

Registration No.: HD 60144431 0001

Report No.: 17063052 007

Manufacturer: Shenzhen BSX Technology
Electronics Co., Ltd.
Rm301.3F&4F 8th Building
LiHao Industrial Area
No. 78 AiNan Road, Longgang
ShenZhen
518116 Guangdong
P.R. China

Products: Blood Pressure Monitors
Infrared Thermometers
Replaces Approval, Registration No.: DD 60133947 0001



Expiry Date: 2023-12-09

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

Effective Date: 2020-03-12

Date: 2020-03-12

Notified Body


Jason Pan


TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

Certificate

The Certification Body of
TÜV Rheinland LGA Products GmbH

hereby certifies that the organization

**Shenzhen BSX Technology
Electronics Co., Ltd.
Rm301.3F&4F 8th Building
LiHao Industrial Area
No. 78 AiNan Road, Longgang
ShenZhen
518116 Guangdong
P.R. China**

has established and applies a quality management system for medical devices
for the following scope:

**Design and Development, Manufacture and Distribution of
Blood Pressure Monitors, Infrared Thermometers**

Proof has been furnished that the requirements specified in

EN ISO 13485:2016

are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2020-03-12
Certificate Registration No.: SX 60144432 0001
An audit was performed. Report No.: 17063052 007
This Certificate is valid until: 2021-12-09

Certification Body



Date 2020-03-12



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
Tel.: +49 221 806-1371 Fax: +49 221 806-3935 e-mail: cert-validity@de.tuv.com <http://www.tuv.com/safety>